

# **EPiC Seating System Adapter Set**

Installation Guide

EPiC Seating System Adapter Tube Set for the **Ottobock Discovery** Wheelchair



#### **Customer Satisfaction**

**Stealth Products** is committed to 100% customer satisfaction. Your complete satisfaction is important to us. Please contact us with feedback or suggestions to help us improve the quality and usability of our products.

You may reach us at:

## **Stealth Products, LLC**

104 John Kelly Drive, Burnet, TX 78611



Phone: (512) 715-9995 Fax: (512) 715-9954

995 Toll Free: (800) 965-9229

info@stealthproducts.com www.stealthproducts.com



PRODUCTS EUROP
De Zwaan 3
1601 MS Enkhuizen
The Netherlands



#### SODIMED SA

Chemin Praz Devant 12 1032 Romanel sur-Lausanne Switzerland **UK** Authorized Representative

# PRIDE MOBILITY PRODUCTS LTD.

32 Wedgwood Road, Bicester, Oxfordshire OX26 4UL United Kingdom

Toll Free: (800) 806-1225

### General

Read and understand all instructions prior to the use of the product. Failure to adhere to instructions and warnings in this document may result in property damage, injury, or death. Product misuse or failure to follow instructions will void the warranty.

Immediately discontinue use if any function is compromised, if parts are missing or loose, or if any component shows signs of excessive wear. Consult with your supplier for repair, adjustment, or replacement.

All persons responsible for fitting, adjustment, and daily use of the devices discussed in these instructions must be familiar with and understand all safety aspects of the devices mentioned. In order for our products to be used successfully, you must read and understand all instructions and warnings, and maintain our products according to our instructions on care and maintenance.

The installation instructions will guide you through this product's options and possibilities.


Instructions are written with the expressed intent of use with standard configurations. They also contain important safety and maintenance information, as well as describe possible problems that can arise during use. For further assistance, or more advanced applications, please contact your supplier or **Stealth Products** at (512) 715-9995 or toll free at (800) 965-9229.

Always keep the operating instructions in a safe place so they may be referenced as necessary.

All information, pictures, illustrations, and specifications are based on the product information that was available at the time of printing. Pictures and illustrations shown in these instructions are representative examples and are not intended to be exact depictions of the various parts of the product.

### Installation Instructions

- 1. Carefully uninstall the levers for brake and tilt functions from the push handles and back support canes, if necessary. Position the cables such that they will not get pinched or kinked during the installation.
- 2. Remove the two fasteners securing the existing mounting plates to the seat tubes on both sides of the wheelchair. Retain the fasteners to use in the next step.
- Select the hole corresponding to the desired seat depth and use the two
  fasteners retained in the previous step to attach the mounting plate of the
  Attachment Hardware to the seat tube.
- 4. Locate the 20" hook and loop strap and thread it through the rings attached to the locking pins on the bottom of the adapter tubes. Leave some slack in the strap and stick the hook and loop surfaces together as shown in the image to the right.



<sup>\*</sup> For questions specific to the frame, please consult the manual included with your Ottobock chair.



### © 2025, Stealth Products, LLC







P147D695-020R2 Sep 12, 2025